


UPDATE INSTRUCTIONS rapidcount X 400


1.

Connect the power cable. Leave the device switched off.




2.

Connect the USB stick to the device (front side).



3.

Now switch on the device.



4.

The update will start automatically.



5.

As soon as the update has been carried out, the messages “Update success” and “Please restart” will appear in the display.



6.

Switch the device off again.

7.

Now remove the USB stick. Your device is now updated.

After each update, we recommend performing a calibration. The procedure can be found in the separate calibration instructions.